



Certificate of Compliance

Certificate: 1671119 (LR 18656)

Master Contract: 158394

Project: 2117696

Date Issued: 2008/12/10

Issued to: Honeywell Inc.
512 Virginia Dr
Fort Washington, PA 19034-3260
USA
Attention: Mr. Edward A. Child

The products listed below are eligible to bear the CSA Mark shown



Issued by: Matthew Brock

Authorized by: Patricia Pasemko, Operations
Manager

PRODUCTS

CLASS 2258 02 - PROCESS CONTROL EQUIPMENT - For Hazardous Locations

Class I, Div. 2, Groups A, B, C, D; Enclosure Type 4X

Universal Dual Analyzer, Model UDA2182, 120/240 VAC, 50/60 Hz, 20VA, 0 to 60 Deg. C ambient, Temperature Code T4; contacts 4A, 120/240 VAC or 30 VDC.

APPLICABLE REQUIREMENTS

CSA C22.2 No. 94-M91 - Special Purpose Enclosures

CSA C22.2 No. 142-M1987 - Process Control Equipment



Certificate: 1671119 (LR 18656)

Master Contract: 158394

Project: 2117696

Date Issued: 2008/12/10

CSA C22.2 No. 213-M1987 - Non-incendive Electrical Equipment for Use in Class I, Division 2 Hazardous Locations

MARKINGS

Submittor's name

Model designation

Electrical input and contact ratings

Hazardous locations designation

Special purpose enclosure designation

Reference to manual

CSA Mark