



EC type-examination Certificate

Number **T10374** revision 0
Project number 11200476
Page 1 of 1

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2004/22/EC, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2004/22/EC, to:

Manufacturer Honeywell Field Solutions
512 Virginia Drive
Fort Washington
Pennsylvania 19034
United States of America

Measuring instrument A model of a **water meter**
Type : VersaFlow Mag 4000 + TWM9000
Environment classes : M1 / E2
Temperature range liquid : +0,1 °C / +50 °C
Temperature range ambient : -25 °C / +55 °C
Measuring range : See § 1.2 of the description

Further properties are described in:
– Description T10374 revision 0;
– Documentation folder T10374-1.

Valid until 15 July 2021

Issuing Authority

NMI Certin B.V., Notified Body number 0122
15 July 2011


C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMI Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMI (see www.nmi.nl).

Reproduction of the complete
document only is permitted.

